

BEMS 3000



In Europe, motor vehicles (light-duty vehicles) will in the future be tested for braking emissions in the WLTP cycle. The basis for this is the directive ECE/TRANS/WP.29/GRPE/2023/4, in short, UN GTR. The particle sizes in brake emissions are in the nanoparticle range of up to about $10 \mu\text{m}$ in concentrations of up to 2×10^6 particles/ cm^3 . Therefore, emissions in this size range are tested for TPN (Total Particle Number, solid and volatile) and SPN (Solid Particle Number, solid particles only, in particles/ cm^3). The $\text{PM}_{2.5}$ and PM_{10} values (in $\mu\text{g}/\text{m}^3$) are also considered.

This device is sold via our partner Link.¹

优势

- Compliance with the new regulations ECE/TRANS/WP.29/GRPE/2023/4
- Integrated flow rate measurement and zero count rate verification
- Measurement paths also available separately for TPN or SPN only
- Monitoring of all data relevant to operation
- Robust, compact design
- Expandable with BEMS 4000 for time-resolved measurement of $\text{PM}_{2.5}$, PM_{10} , TSP, and particle size distribution.

应用领域

- Measurement of brake dust emissions according to UNGTR
- Measurement of the number concentration up to $2.5 \mu\text{m}$ in other applications such as tire wear measurement

¹Link Website: <https://www.linkeng.com/product/model-4222-brake-emissions-particle-measuring-system/>

技术数据

测量原理	Kondensationspartikelzähler
测量范围(粒径)	10–2.500 nm
颗粒物最大数量浓度	0,1–1.000.000 Partikel/cm ³ , single count mode inkl. Verdünnung 1:100, photometrisch -10 ⁸
体积流量	2*5 l/m (Aerosol)
Volume flow (clean air)	180 l/min
接口	Ethernet (LAN)
Protocols	RJ45 / TCP/IP
电源	100/230 V, 50/60 Hz, max. 600 W
Power consumption	Max. 600 W
Installation conditions	Temperaturbereich: +15°C–25°C; Luftfeuchtigkeit: ≤85%; Betriebsdruck am Aerosoleinlass: 850–1.050 mbar absolut
Compressed air supply	4–8 bar ISO, erforderliche Druckluftqualität nach ISO 8573-1:2010, mind. Reinheitsklasse 2
Dilution factor	1:10 / 1:10
Dimensions	1,100 • 750 • 650 mm (H • W • D)
重量	Approx. 135 kg

标准和证书

ECE/TRANS/WP.29/GRPE/2023/4