

# LFT 2000



ISO 29463-1 requires for all HEPA /ULPA filters with an efficiency of 99.95% or higher (ISO 35H/H13) the following test to ensure filter quality:

- local efficiency according to ISO 29463-4
- overall efficiency according to ISO 29463-5

The LFT 2000 combines the requirements of ISO 29463-4 Annex F for Class 35 H filters with a simple manual test procedure.

The system manually detects possible leakage points easily and quickly with the appropriate software.

The output of the test report is for local filter efficiency as well as for leak detection.

## 优势

- Quick manual scanning with OPC
- Test report
- Clear leak detection
- Easy installation of filter elements
- Adapters for various filter dimensions
- Upgradable to an automatic scanning system
- Use of components for ISO 14644 In-Situ integrity testing

## 应用领域

- Classification of HEPA/ULPA filters
- Filter test comparable to ISO 29463-4 Annex F

## 特点

- Selectable particle counters, easily adaptable
- Can be upgraded with automatic scan system
- Individual adaptation of the required test volume flows

## 技术数据

体积流量	100–1.200 m <sup>3</sup> /h	电源	400 V, 50 Hz
Differential pressure measurement	Up to 1,200 Pa	Compressed air supply	6 bar
Size filter element	300 x 300–600 x 1,200 mm		

## 标准和证书

ISO 29463-4, ISO 29463-5